

Feline calicivirus Antigen (FCV-Ag) Rapid Quantitative Test

INTENDED USE

The feline herpes calicivirus antigen (FCV-Ag) Rapid Quantitative Test is a fluorescence immunoassay used along with FIA 680 for quantitative determination of FCV-Ag concentration in feline nasopharyngeal or ocular secretion. The test is used as an aid to infection inspection, the inflammatory response.

For in vitro diagnostic use only. For professional use only.

TEST PRINCIPLE

- 1.This test employs a quantitative Double antibody sandwich fluorescence immunoassay technique.
- 2.The fluorescent signal intensity reflects the amount of FCV-Ag captured and is processed in FIA 680. The FCV-Ag concentration is expressed in Tu/ml.

WARNINGS AND PRECAUTIONS

- 1.This kit is for in vitro diagnostic use only. Do not swallow.
- 2.Lot Number of all the test components (test device, ID chip and buffer) must match each other. Do not mix components from different kit lots.
- 3.Inspect the packaging and labels before use. Do not use if the pouch is broken, torn or not well sealed, or the vial looks damaged or leaked.
- 4.Carefully follow the instructions and procedures described in this insert.
- 5.Do not use the test device beyond the expiration date. The test device must remain in its original sealed pouch until ready to use. Do not use if the pouch or the device itself is damaged, torn or not fully sealed.
- 6.A buffer tube should be used for processing one sample only.
- 7.The operation shall be conducted away from vibration and magnetic field. FIA 680 may generate minute vibration during use, which should be regarded as normal.
- 8.One pipette tip should be used for one specimen only.
- 9.Do not touch the test area of the test device.
- 10.All specimens and used test materials are considered as potentially infectious. The used pipette tips, buffer tubes, test devices and specimens must be handled carefully and disposed of in accordance with local regulations and procedures.

MATERIAL

Material Provided

Each box contains:

- 1.10 individual sealed pouches, each containing:
 - a test Device
 - a desiccant pouch
- 2.One Test Device ID Chip
- 3.Instructions for Use
- 4.10 tubes of Tris-HCl buffer
- 5.10 cotton swabs

Material Required But Not Provided

1. FIA 680
- 2.Transfer Pipette

- 3.Timer
- 4.Centrifuge

STORAGE AND STABILITY

- 1.Store the test kit at 4~30℃ up to the expiration date.
- 2.Once the pouch is opened, the test should be performed within an hour.
- 3.If removed from refrigerator, allow the test for 30 minutes to attain room temperature before testing.

SPECIMEN COLLECTION AND PREPARATION

The test can be performed with either nasopharyngeal or ocular swab. Use a cotton swab to take samples from the eye, nose or mouth of feline. Place the swab with the sample in buffers. Stir thoroughly and centrifuge for 1 minute.

Optimally, the test should be performed immediately after the specimen collection. If the test cannot be performed within 2 hour after specimen collection, store the specimen at 2℃~8℃ for no longer than 48 hours. For long-term storage, specimens shall be kept below -20℃.

Bring all materials to room temperature before use. Frozen specimens must be completely thawed and mixed well prior to testing. Specimens should not be frozen and thawed repeatedly. Only clear, non-hemolyzed specimens can be used.

TEST PROCEDURE

Refer to FIA 680 Operation Manual for complete instructions for use of the Test device.

- 1.Set a Test Device on a clean, level horizontal place.
- 2.Make sure that the Test Device lot No. matches with ID Chip No. Insert ID Chip into the meter. Be aware not to touch the insertion tip of the ID chip.
- 3.Pipette 100 µl of mixed sample to add into the sample of the test device. Avoid forming bubbles.
- 4.Please refer to the Section V Operation in Operation Manual for details.
 - a) **Quick Test mode:** start the timer right after adding the sample mixture to the sample well. Leave the Test Device at room temperature for 10 minutes. Then insert test device immediately onto the holder of the meter and click Test. The instrument will scan the Test Device automatically and show the test result.
 - b) **Standard Test mode:** insert the Test Device into the device holder of the Meter right after adding the sample to the sample well, click Test. The meter will start to countdown and read the test result automatically.
- 6.Results are displayed on the main screen or be printed by click Print.

INTERPRETATION OF RESULTS

- 1.The FIA 680 calculates FCV-Ag test results automatically and displays the concentration of FCV-Ag on the screen right after correctly adding the sample to the sample well for 3 minutes. For further information, refer to the Operation Manual for the FIA 680.
- 2.FIA 680 will prompt "No Sample or sample volume insufficient!" when insufficient sample volume or liquid is not fully crawled across the test line, then recommends adding sample again.

QUALITY CONTROL

Each FCV-Ag Rapid Quantitative Test contains an internal control for routine quality control requirements. This internal quality control is performed each time a patient sample is tested, If an invalid result from the internal control occurs, the meter will display an error message, indicating that another test

should be conducted.

LIMITATIONS OF PROCEDURE

- 1.This test is developed for feline nasopharyngeal or ocular swab specimen.
- 2.The results of FCV-Ag Rapid Quantitative Test should be evaluated with all clinical and laboratory data available. If test results do not agree with the clinical evaluation, additional tests should be performed accordingly.
- 3.The false positive results include cross-reactions with some components of the nasopharyngeal or ocular secretion from individual to antibodies; and non-specific adhesion of some components in feline nasopharyngeal or ocular secretion. that have similar epitopes to capture and detector antibodies. In the case of false negative results, the most common factors are: non-responsiveness of antigen to the antibodies by that certain unknown components are masking its epitope, such that antigen cannot be seen by the antibodies; instability of FCV-Ag antigen, resulting in degradation with time and, or temperature, such that they become no longer recognizable by antibodies; and degraded other test components. The performance of the test is highly sensitive to the storage and handling conditions of kits and sample specimens at optimal conditions.
- 4.There is the possibility that factors such as technical or procedural errors, as well as additional substances in feline nasopharyngeal or ocular secretion specimens that are not listed in the Cross-Reactivity section, may interfere the test and cause erroneous results.

INTERPRETATION

Reference range of FCV-Ag in feline nasopharyngeal or ocular secretion.

- 1.Detection range: 5-200 Tu/ml
 - 2.Reference: 10 Tu/ml
- Each Laboratory should establish a reference range that is representative of the population to be evaluated.

MANUFACTURED BY

Guangzhou Ditron Medtech Co., Ltd
Email: info@ditronmed.com
Address: No.06, Block B16, Fangcun Avenue East, Liwan District, Guangzhou, P.R.China